



Consolidated Laboratories, Inc.

732 Arrow Grand Circle
Covina, California 91722
(626) 915-8991
FAX (626) 966-3156

REPORT NO. 11858

PAGE 2

TEST REPORT

4. **TEST EQUIPMENT** (Continued)

- Hydraulic Boost Pump: SC Hydraulic Engineering Model SC10-500-25, S/N 54504, at air drive pressures up to 100 PSIG with hydraulic outlet pressures up to 40,000 PSIG. Calib. N/A.
- Hydraulic Pump (#11): Pacific Pumping Company (PACO) Cat.# 10-12705-130001-1731, S/N B2HF22531C, with 7.5 HP, 3500 RPM drive motor. Variable flow and pressure to 118 GPM at 90 PSIG. Calib. N/A.
- Pressure Gauge: Ashcroft 6" Maxisafe Test Gauge, ID/N 105, 0 - 6000 x 20 PSIG, $\pm 0.5\%$. Calib. due 3-8-03, 6 mos.
- Pressure Gauge: U.S. 6" Test Gauge, ID/N 108, 0 - 160 x 1 PSIG, $\pm 0.25\%$. Calib. due 3-13-03, 6 mos.
- Pressure Gauge: Barton 6" Differential gauge, ID/N 139, 0 - 20 x 0.2 PSID, $\pm 1.0\%$. Calib. due 7-6-03, 6 mos.
- Pressure/Vacuum Gauge: 3D Instruments 4½" Test Gauge, ID/N 234, 0 - 100 x 1 PSIG, $\pm 5\%$ and 0 - 30 x 2 In. Hg., $\pm 1\%$. Calib. due 3-13-03, 6 mos.

